



*High Reliability &
Automatization
With
Convenience*



WAP LAB U-500
Automated Urine Analyzer

Global Diagnostics Supplier

WAP LAB U-500

Automated Urine Analyzer

FDA 510 (K) NO. K082811 Approved

| SPECIFICATIONS |

Parameters : Leukocyte, ketone, Nitrite, Urobilinogen, Bilirubin,
Protein, Glucose, Specific Gravity, blood, pH, Ascorbic Acid,
Microalbumin, Creatinine, Calcium

Measurement Principle : Dual Wavelength Reflectance Photometry

Reaction Time : 65 Seconds

Throughput : 520 Samples / Hour

Display : LCD Touch Screen

Memory Capacity : 999 Results

Strip : URIT 10G/11G/14G Urine Reagent Strip.

Interface : Serial Port (RS232), Parallel Port (For External Printer),
PS/2 (For Barcode Reader), USB Port (For Data Output)

Work Condition : Environmental Temperature : 15-30°C

RH: 20% - 80%, Luminance : = 2000 Lx

Power Supply : 100-240 AC, 20/60hz

Dimension : 390 mm X 340 mm X 290 mm (Depth X Width X Height)

Weight : 6.5 Kg



| FEATURES & FUNCTION |

- Friendly Operation With Touch Screen
- Easy Maintenance Design With Auto Waste Collection
- High Test Speed, Idea For Extensive Sample Laboratory
- Fully Automatic Operation For Ready - To - Test
- Internal Calibration For Sensitivity Of Strips
- RS232 SERIAL INTERFACE TO HOST PC/LIS, Barcode Reader Optional
- URIT UQ-10, UQ-11 AND UQ-14 Urinalysis Control

Manufactured By :
Urit Medical Co., Ltd.
No. D-07 Information Industry
High - Tech Zone, Guilin, Guangxi 541004
P. R. China
Fax: + 86 - 773 - 2288560

